



***** PARAMEDIC ALERT *****

No. 2011-06

Date: July 25, 2011

Topic: ZOFRAN RECALL

Effective: Immediately

Attention: All EDC ALS Service Agencies

Hospira, Inc. is voluntarily recalling the iSecure Syringe products identified below to address complaints about cracks in some lots of the syringes. These cracks, which have only been found in the original iSecure syringes with the white plunger rod, have resulted in low filled, unfilled or leaking products and present the potential for contamination of the solution. iSecure products with the purple plunger rod are not affected.

Life-Assist Item#: DR1120-62

Ondansetron, 4 mg, 2 ml iSecure Luer Lock Syringe

NDC#: 0409-1120-62

Manufacturer: Hospira, Inc.

Lot Numbers: 01695LL; 03820LL; 79610LL; 82610LL; 84515LL; 86665LL;
88585LL; 94795LL; 95640LL

In the meantime, all service agencies should utilize either oral dissolving tablets or Zofran supplied in vials until preloads become available again. You should contact your suppliers for return and replacement of any recalled medication.

Sincerely,

A handwritten signature in black ink, appearing to read "David Brazzel", with a stylized flourish at the end.

David Brazzel, MD

Medical Director, El Dorado County EMS Agency